

# EC Declaration of Conformity

<b>Manufacturer/ Supplier Information:</b>	<b>BioFire Diagnostics, LLC</b> 515 Colorow Drive Salt Lake City, Utah 84108, USA Phone: 1-801-736-6354 <a href="mailto:regulatory@BioFireDX.com">regulatory@BioFireDX.com</a> <a href="http://www.BioFireDX.com">http://www.BioFireDX.com</a>
--	---

We, BioFire Diagnostics, LLC, declare under our sole responsibility, that the product

## FilmArray® Pneumonia Panel (RFIT-ASY-0144, RFIT-ASY-0145)

meets the provisions of the European Directive 98/79/EC for *in vitro* Diagnostic Medical Devices. The device is classified as an *In Vitro* Diagnostic (IVD) Device under Annex II list B. BioFire Diagnostics' quality system is registered to EN ISO 13485:2016.

The following relevant standards have been met:

<b>EN ISO 13485:2016</b> Medical devices – Quality Management System – Requirements for regulatory purposes
<b>EN ISO 14971:2012</b> Medical devices – Application of risk management to medical devices
<b>EN 13641:2002</b> Elimination or reduction of risk of infection related to <i>in vitro</i> diagnostic reagents
<b>EN 62366:2008</b> Medical devices-Application of usability engineering to medical devices
<b>EN 13612:2002</b> Performance evaluation of <i>in vitro</i> diagnostic medical devices
<b>EN ISO 23640:2015</b> <i>In vitro</i> diagnostic medical devices – Evaluation of stability of <i>in vitro</i> diagnostic reagents
<b>EN ISO 15223-1:2016</b> Medical Devices – Symbols to be used with medical device labels, labeling and information to be supplied – Part 1: General requirements
<b>EN ISO 18113-1:2011</b> <i>In vitro</i> diagnostic medical devices – Information supplied by the manufacturer (labeling) – Part 1: Terms, definition and general requirements
<b>EN ISO 18113-2:2011</b> <i>In vitro</i> diagnostic medical devices – Information supplied by the manufacturer (labeling) – Part 2: <i>In vitro</i> diagnostic reagents for professional use

Technical documentation demonstrating compliance as described in Annex IV of the European Directive 98/79/EC is kept by the manufacturer and can be made available by the authorized representative in Europe - QARAD bvba, Ciplastraat 3, B-2440 Geel, Belgium.

The notified body for this product is BSI (Notified body #0086; Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes, MK5 8PP, United Kingdom).

Salt Lake City, UT, USA 11-14-18  
(Place and date of issue)

  
**Randy Rasmussen**  
Chief Executive Officer

**BIO FIRE** • 515 Colorow Drive, Salt Lake City, UT 84108  
phone 1-801-736-6354 | fax 1-801-588-0507  
[www.BioFireDX.com](http://www.BioFireDX.com)

A BIOMÉRIEUX COMPANY